

CERTIFICATE OF ANALYSIS

page 1 of 1

Sigma-Aldrich Laborchemikalien GmbH D-30918 Seelze

Telefon: +49 5137 8238-0

07.03.2012/329896

Analysed for
Sigma-Aldrich Laborchemikalien GmbH
Production HYDRANAL®
Wunstorfer Str. 40
D-30926 Seelze

QC Release Date: 05.Mar.2012

Production Date: 27.Feb.2012

Expiry Date: 11.Feb.2015

Article/Product: 34801**Lot : SZBC0580V****HYDRANAL-Titrant 5**

The quality certificate is valid for the time of delivery

HYDRANAL-Titrant 5 is a volumetric titrant for the water determination according to the Karl Fischer method. It is used in conjunction with HYDRANAL-Solvent.

A Suitability Test was carried out on this Lot after production.

The titre is determined according to ISO 760 and ASTM E 203.

In this Lot we found a titre of 5.02 mg/ml.

- The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

**Sigma-Aldrich Laborchemikalien GmbH
Thomas Wendt
Supervisor Technical Service HYDRANAL
Wunstorfer Str. 40, D-30926 Seelze**



DGA-PL-6670.09