

CERTIFICATE OF ANALYSIS
page 1 of 1

Sigma-Aldrich Laborchemikalien GmbH D-30918 Seelze
Telefon: +49 5137 8238-0

22.05.2012/372822

Analysed for
Sigma-Aldrich Laborchemikalien GmbH
Production HYDRANAL®
Wunstorfer Str. 40
D-30926 Seelze

QC Release Date: 18.May.2012
Production Date: 14.May.2012
Expiry Date: 29.Apr.2015

Article/Product: 34805

Lot : SZBC1350V

HYDRANAL-Composite 5

The quality certificate is valid for the time of delivery

HYDRANAL-Composite 5 is a titrating agent for the volumetric Karl Fischer titration.

A Suitability Test was carried out on this Lot after production.


The titre is determined according to ISO 760 and ASTM E 203.

In this Lot we found a titre of 5.37 mg water per ml.

The titre decreases slightly during storage.

- The minimum shelf life is based on the current knowledge and holds only for proper storage conditions in the originally closed flasks/ packages.
- We herewith confirm that the delivery is effected according to the technical delivery conditions agreed.
- Particular properties of the products or the suitability for a particular area of application are not assured.
- We guarantee a proper quality within our General Conditions of Sales.

Sigma-Aldrich Laborchemikalien GmbH
Thomas Wendt
Supervisor Technical Service HYDRANAL
Wunstorfer Str. 40, D-30926 Seelze



DGA-PL-6670.09